



Waveland @ Randall Medical Building

2nd Floor; Tag# 102

Last AED Status 12 : 27 : 05 pm on 03/10/2022

Admin

Account Manager

Make / Model

Philips -HeartStart FR2

Serial Number

0306186423

Click the icon that matches the status of your AED.



Ready (Blinking Black Hourglass)



Not Ready (Red X). Service Immediately and Call Premedics

Oxygen Included

No

Oxygen Status

<input type="checkbox"/>	<input type="checkbox"/>
<input type="checkbox"/>	<input type="checkbox"/>

Green Yellow Red

Your AED status **READY**

Additional Equipment

Description	Expiration	Install By	Lot #	Notes	Reference #	Approximate Install Date
Primary Battery	01/17/2023	09/01/2024	79130P	Sync HW1 Unit.		01/17/2019
Backup Battery						
Primary Adult Pads	01/31/2023		00709-4035	SH 10/2/21; exp 1/31/23; 00703-4034	989803149981	10/01/2021
Backup Adult Pads						
Child Pads	07/31/2022	Reducer/Key :	Y011520-01	SH 10/2/21; exp 7/31/22; Y011520-01	M3870A	2021-10-01

Other Equipment

Response Pack : Yes	Trauma Bag : Bag :	Good/Visible Signage : Yes	Cabinet/Wall mount : Yes	AED Self-test : Yes	AED Cabinet Alarm Test : Yes
------------------------	-----------------------	-------------------------------	-----------------------------	------------------------	---------------------------------

Software	2005
----------	------

Out of Date	
G2010 Compliant	Yes