



**Waveland @ Randall Medical Building**

**2nd Floor; Tag# 102**

Last AED Status 09 : 14 : 13 pm on 03/03/2021

**Admin**

Chad Francis

**Make / Model**

Philips -HeartStart FR2

**Serial Number**

0306186423

Click the icon that matches the status of your AED.



Ready (Blinking Black Hourglass)



Not Ready (Red X). Service Immediately and Call Premedics

**Oxygen Included**

No

**Oxygen Status**

<input type="checkbox"/>	<input type="checkbox"/>
<input type="checkbox"/>	<input type="checkbox"/>

Green Yellow Red

Your AED status **READY**

**Additional Equipment**

Description	Expiration	Install By	Lot #	Notes	Reference #	Approximate Install Date
Primary Battery	01/17/2023	09/01/2024	79130P	Sync HW1 Unit.		01/17/2019
Backup Battery						
Primary Adult Pads	12/31/2021		70724-1805	sh 1-3-20; exp 10-31-21; 90409-4039		
Backup Adult Pads						
Child Pads	11/30/2021	Reducer/Key :	Y110317-03	sh 1-3-20; exp 11-30-21; Y050219-10	M3870A	2018-05-02

**Other Equipment**

Response Pack : Yes	Trauma Bag : Bag :	Good/Visible Signage : Yes	Cabinet/Wall mount : Yes	AED Self-test : Yes	AED Cabinet Alarm Test : Yes
------------------------	-----------------------	-------------------------------	-----------------------------	------------------------	---------------------------------

Software	2005
----------	------

Out of Date	
G2010 Compliant	Yes